



**ADDENDUM NUMBER:
November 4, 2021**

For

**IFB 742-22-192-1
AEDs**

**University of Texas Permian Basin
4901 E. University Blvd.
Odessa, TX 79762**

The following revisions, additions and clarifications shall be incorporated in the Request for Proposal referenced above. All other provisions of the Request for Proposal shall remain unchanged.

Question and Answers

- 1) If the AED is going into a wall cabinet, we generally recommend the “Handle” version of the CR2 instead of the “Case” version. Without the Case, the AED is simpler, lighter/smaller, and less expensive. (We recommend the case version for AEDs going into vehicles.
We prefer the case model for all units as most will be in a vehicle.

- 2) We also generally deploy the USB version without WiFi, as this model is less expensive, and managing WiFi connectivity across large deployments can be challenging. State law requires that AEDs be checked, in person, once each month, so WiFi connectivity may also provide less benefit than envisioned.
We prefer the wifi for a few different solutions.

3. We generally deploy Fully-Automatic instead of Semi-Automatic, as this removes one potential point of failure. The Fully-Automatic AED is not much more expensive.
We train and for the use of semi-automatic and prefer it for several reasons.

4. Are you able to provide a date or estimate of when you will need items delivered by, once contract is awarded? There is reference to 14 days being the default delivery timeframe. Will you be placing one order or is there a certain contract period within which you will be placing multiple orders?
**We will be happy to accept a partial order in December first week and the remainder before the end of January.
If this is not possible due to the supply chain problem, we can accept three shipments in December, January, and February.**

The following revisions, additions and clarifications shall be incorporated in the Request for Competitive Sealed Proposal referenced above. All other provisions of the Request for Proposal shall remain unchanged. BIDDERS MUST ACKNOWLEDGE RECEIPT OF ADDENDUM(S)

signature _____

Pam Jenkins, Procurement Specialist UTPB

End of Addendum